

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01724	Calib. Date: 09/21/2021
Certification File No.:	01725	Calib. No.: 00055
Linearity File No.:	01726	Cert. Date: 09/21/2021
Solution File No.:	01734	Cert. No.: 00039
Sequential File No.:	01734	Lin. Date: 09/21/2021
		Lin. No.: 00038
		Soln. Date: 10/19/2021
		Soln. No.: 00293
		File Date: 10/19/2021

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	07/06/2022
Solution Control Lot:	20320			Bottle No.:	1468

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:53D		
Control 1 EC	0.099%	20:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:54D		
Control 2 EC	0.098%	20:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:55D		
Control 3 EC	0.098%	20:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:57D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: PH  #97

Badge No.: 97

Date: 10/19/2021