

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01694	Calib. Date:	03/31/2021
Certification File No.:	01695	Cert. No.:	00054
Linearity File No.:	01696	Cert. Date:	03/31/2021
Solution File No.:	01697	Lin. Date:	03/31/2021
Sequential File No.:	01697	Lin. No.:	00037
		Soln. Date:	03/31/2021
		Soln. No.:	00285
		File Date:	03/31/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0364
Solution Control Lot:	20320	Expires:	07/06/2022
		Bottle No.:	1452

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:12S	34.0°C	
Control 1 EC	0.100%	12:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13S	34.0°C	
Control 2 EC	0.100%	12:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14S	34.0°C	
Control 3 EC	0.100%	12:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:15S	34.0°C	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist

Temperature Probe Serial Number: DDWAP2-215

MRW

Changed By:

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature: *Matthew Watson*

Matthew Watson #7078

Badge No.: 7078

Date: 03/31/2021