

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01694 Calib. Date: 03/31/2021 Calib. No.: 00054
Certification File No.: 01695 Cert. Date: 03/31/2021 Cert. No.: 00038
Linearity File No.: 01696 Lin. Date: 03/31/2021 Lin. No.: 00037
Solution File No.: 01704 Soln. Date: 05/27/2021 Soln. No.: 00287
Sequential File No.: 01704 File Date: 05/27/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 1364

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:15D		
Control 1 EC	0.097%	15:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:16D		
Control 2 EC	0.097%	15:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17D		
Control 3 EC	0.098%	15:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: _____



Badge No.: 97

Date: 05/27/2021