

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE	Calib. No.:	00057
Calibration File No.:	01770	Cert. Date:	07/12/2022
Certification File No.:	01771	Lin. Date:	07/12/2022
Linearity File No.:	01772	Soln. Date:	10/03/2022
Solution File No.:	01783	File Date:	10/03/2022
Sequential File No.:	01783		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	09/01/2023
Solution Control Lot:	21360			Bottle No.:	1403

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:05D		
Control 1 EC	0.098%	13:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:06D		
Control 2 EC	0.097%	13:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:07D		
Control 3 EC	0.097%	13:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: PH. [Signature] # 585

Badge No.: 585

Date: 10/03/2022