

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01770	Calib. Date: 07/12/2022
Certification File No.:	01771	Calib. No.: 00057
Linearity File No.:	01772	Cert. Date: 07/12/2022
Solution File No.:	01775	Cert. No.: 00041
Sequential File No.:	01775	Lin. Date: 07/12/2022
		Lin. No.: 00040
		Soln. Date: 08/07/2022
		Soln. No.: 00307
		File Date: 08/07/2022

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1306

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:31D		
Control 1 EC	0.099%	18:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:32D		
Control 2 EC	0.098%	18:33D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:34D		
Control 3 EC	0.098%	18:34D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: _____



Badge No.: 089

Date: 08/07/2022