

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01830 Calib. Date: 06/06/2023 Calib. No.: 00059
Certification File No.: 01831 Cert. Date: 06/06/2023 Cert. No.: 00043
Linearity File No.: 01832 Lin. Date: 06/06/2023 Lin. No.: 00042
Solution File No.: 01850 Soln. Date: 10/11/2023 Soln. No.: 00324
Sequential File No.: 01850 File Date: 10/11/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363
Control Solution %: 0.100% Expires: 02/01/2025
Solution Control Lot: 23040 Bottle No.: 1051

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:06D		
Control 1 EC	0.097%	08:06D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:07D		
Control 2 EC	0.097%	08:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:08D		
Control 3 EC	0.096%	08:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: Edward J McQuade #97

Badge No.: 97

Date: 10/11/2023