

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
Location: WEST WINDSOR POLICE  
Calibration File No.: 01830 Calib. Date: 06/06/2023 Calib. No.: 00059  
Certification File No.: 01831 Cert. Date: 06/06/2023 Cert. No.: 00043  
Linearity File No.: 01832 Lin. Date: 06/06/2023 Lin. No.: 00042  
Solution File No.: 01837 Soln. Date: 08/02/2023 Soln. No.: 00321  
Sequential File No.: 01837 File Date: 08/02/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363  
Control Solution %: 0.100% Expires: 09/01/2023  
Solution Control Lot: 21360 Bottle No.: 1305

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:53D		
Control 1 EC	0.098%	12:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		
Control 2 EC	0.097%	12:54D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55D		
Control 3 EC	0.097%	12:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: \_\_\_\_\_

*Edward J. McQuade*

Badge No.: 97

Date: 08/02/2023