

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01118	Calib. Date: 07/09/2014
Certification File No.:	01119	Calib. No.: 00032
Linearity File No.:	01120	Cert. Date: 07/09/2014
Solution File No.:	01154	Cert. No.: 00020
Sequential File No.:	01154	Lin. Date: 07/09/2014
		Lin. No.: 00020
		Soln. Date: 11/20/2014
		Soln. No.: 00164
		File Date: 11/20/2014

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %:	0.100%		Expires: 07/15/2015
Solution Control Lot:	13G115		Bottle No.: 0007

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:48S		
Control 1 EC	0.100%	15:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:50S		
Control 2 EC	0.099%	15:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51S		
Control 3 EC	0.099%	15:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BROWN

First Name: KYLE

MI: W

Signature: \_\_\_\_\_



Badge No.: 89

Date: 11/20/2014