

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01030	Calib. Date: 08/14/2013
Certification File No.:	01031	Calib. No.: 00030
Linearity File No.:	01032	Cert. Date: 08/14/2013
Solution File No.:	01062	Cert. No.: 00018
Sequential File No.:	01062	Lin. Date: 08/14/2013
		Lin. No.: 00018
		Soln. Date: 12/20/2013
		Soln. No.: 00146
		File Date: 12/20/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	12L109	Expires: 12/26/2014
		Bottle No.: 0119

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:16S		
Control 1 EC	0.100%	10:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17S		
Control 2 EC	0.098%	10:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19S		
Control 3 EC	0.099%	10:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: MARYLOUIS

MI: -

Signature: 

Badge No.: 431

Date: 12/20/2013