

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00984	Calib. Date:	02/26/2013
Certification File No.:	00985	Cert. Date:	02/26/2013
Linearity File No.:	00986	Lin. Date:	02/26/2013
Solution File No.:	00994	Soln. Date:	03/24/2013
Sequential File No.:	00994	File Date:	03/24/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	12C100	Expires:	03/02/2014
		Bottle No.:	0423

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:14S		
Control 1 EC	0.099%	10:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16S		
Control 2 EC	0.098%	10:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17S		
Control 3 EC	0.098%	10:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: EVANS

First Name: LEE

MI: F

Badge No.: 430

Date: 03/24/2013

Signature: \_\_\_\_\_

Sgt Lee Evans #430

7