

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
Location: WEST WINDSOR POLICE  
Calibration File No.: 01169 Calib. Date: 01/08/2015 Calib. No.: 00033  
Certification File No.: 01170 Cert. Date: 01/08/2015 Cert. No.: 00021  
Linearity File No.: 01171 Lin. Date: 01/08/2015 Lin. No.: 00021  
Solution File No.: 01210 Soln. Date: 04/16/2015 Soln. No.: 00174  
Sequential File No.: 01210 File Date: 04/16/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363  
Control Solution %: 0.100% Expires: 08/22/2016  
Solution Control Lot: 14H132 Bottle No.: 1272

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:30D		
Control 1 EC	0.101%	11:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32D		
Control 2 EC	0.099%	11:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33D		
Control 3 EC	0.099%	11:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BROWN

First Name: KYLE

MI: W

Signature: \_\_\_\_\_



Badge No.: 89

Date: 04/16/2015