

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00984	Calib. Date:	02/26/2013
Certification File No.:	00985	Cert. Date:	02/26/2013
Linearity File No.:	00986	Lin. Date:	02/26/2013
Solution File No.:	01014	Soln. Date:	05/28/2013
Sequential File No.:	01014	File Date:	05/28/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	12L109	Expires:	12/26/2014
		Bottle No.:	0107

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:38D		
Control 1 EC	0.101%	14:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:39D		
Control 2 EC	0.100%	14:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:40D		
Control 3 EC	0.099%	14:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: _____

Kyle Brown #89

Badge No.: 589

Date: 05/28/2013