

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00984	Calib. Date: 02/26/2013
Certification File No.:	00985	Calib. No.: 00029
Linearity File No.:	00986	Cert. Date: 02/26/2013
Solution File No.:	01024	Lin. Date: 02/26/2013
Sequential File No.:	01024	Lin. No.: 00017
		Soln. Date: 07/05/2013
		Soln. No.: 00138
		File Date: 07/05/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	12L109	Expires: 12/26/2014
		Bottle No.: 0106

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:38D		
Control 1 EC	0.100%	10:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39D		
Control 2 EC	0.099%	10:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:40D		
Control 3 EC	0.099%	10:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 80

Date: 07/05/2013