

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01377	Calib. Date:	09/22/2016
Certification File No.:	01378	Cert. No.:	00038
Linearity File No.:	01379	Cert. Date:	09/22/2016
Solution File No.:	01403	Lin. No.:	00025
Sequential File No.:	01403	Lin. Date:	09/22/2016
		Soln. Date:	01/05/2017
		Soln. No.:	00207
		File Date:	01/05/2017

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	05/17/2018
Solution Control Lot:	201605D			Bottle No.:	0483

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:58S		
Control 1 EC	0.103%	07:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:59S		
Control 2 EC	0.102%	08:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	08:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01S		
Control 3 EC	0.102%	08:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	08:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____

Francisco Latorre #68

Badge No.: 68

Date: 01/05/2017