

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01316	Calib. Date: 03/28/2016
Certification File No.:	01317	Calib. No.: 00037
Linearity File No.:	01318	Cert. Date: 03/28/2016
Solution File No.:	01374	Cert. No.: 00024
Sequential File No.:	01374	Lin. Date: 03/28/2016
		Lin. No.: 00024
		Soln. Date: 09/15/2016
		Soln. No.: 00201
		File Date: 09/15/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	201603L	Expires: 03/22/2018
		Bottle No.: 0867

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:12D		
Control 1 EC	0.099%	09:12D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13D		
Control 2 EC	0.099%	09:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14D		
Control 3 EC	0.099%	09:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MOHR

First Name: DANNY

MI:

Signature: _____



Badge No.: 471

Date: 09/15/2016