

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01377	Calib. Date:	09/22/2016
Certification File No.:	01378	Cert. No.:	00038
Linearity File No.:	01379	Cert. Date:	09/22/2016
Solution File No.:	01383	Lin. No.:	00025
Sequential File No.:	01383	Lin. Date:	09/22/2016
		Soln. Date:	10/14/2016
		Soln. No.:	00203
		File Date:	10/14/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	05/17/2018
Solution Control Lot:	201605D			Bottle No.:	0485

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:06D		
Control 1 EC	0.102%	07:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	07:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:07D		
Control 2 EC	0.102%	07:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	07:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:08D		
Control 3 EC	0.102%	07:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	07:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MOHR

First Name: DANNY

MI:

Signature: _____



Badge No.: 471

Date: 10/14/2016