

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01222	Calib. Date: 06/10/2015
Certification File No.:	01223	Calib. No.: 00034
Linearity File No.:	01224	Cert. Date: 06/10/2015
Solution File No.:	01259	Cert. No.: 00022
Sequential File No.:	01259	Lin. Date: 06/10/2015
		Lin. No.: 00022
		Soln. Date: 10/03/2015
		Soln. No.: 00183
		File Date: 10/03/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 04/13/2017
Solution Control Lot:	15D138		Bottle No.: 0175

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:16D		
Control 1 EC	0.099%	10:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17D		
Control 2 EC	0.098%	10:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		
Control 3 EC	0.099%	10:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BAL

First Name: FRANK

MI: N

Signature: _____



Badge No.: 567

Date: 10/03/2015