

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWF-0382 |
| Location: | WEST WINDSOR POLICE | |
| Calibration File No.: | 01377 | Calib. Date: 09/22/2016 |
| Certification File No.: | 01378 | Calib. No.: 00038 |
| Linearity File No.: | 01379 | Cert. Date: 09/22/2016 |
| Solution File No.: | 01387 | Cert. No.: 00025 |
| Sequential File No.: | 01387 | Lin. Date: 09/22/2016 |
| | | Lin. No.: 00025 |
| | | Soln. Date: 11/09/2016 |
| | | Soln. No.: 00204 |
| | | File Date: 11/09/2016 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWJ S3-0363 |
| Solution Control Lot: | 201605D | Expires: 05/17/2018 |
| | | Bottle No.: 0486 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 12:35S | | |
| Control 1 EC | 0.103% | 12:35S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.103% | 12:35S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:36S | | |
| Control 2 EC | 0.103% | 12:37S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.103% | 12:37S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:37S | | |
| Control 3 EC | 0.102% | 12:38S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.104% | 12:38S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:39S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____



Badge No.: 68

Date: 11/09/2016