

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01604	Calib. Date: 07/29/2019
Certification File No.:	01605	Calib. No.: 00046
Linearity File No.:	01606	Cert. Date: 07/29/2019
Solution File No.:	01623	Cert. No.: 00033
Sequential File No.:	01623	Lin. Date: 07/29/2019
		Lin. No.: 00032
		Soln. Date: 11/26/2019
		Soln. No.: 00258
		File Date: 11/26/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 0102

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:45S		
Control 1 EC	0.098%	08:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46S		
Control 2 EC	0.097%	08:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48S		
Control 3 EC	0.096%	08:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature:  #585

Badge No.: 585

Date: 11/26/2019