

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01222	Calib. Date: 06/10/2015
Certification File No.:	01223	Calib. No.: 00034
Linearity File No.:	01224	Cert. Date: 06/10/2015
Solution File No.:	01237	Cert. No.: 00022
Sequential File No.:	01237	Lin. Date: 06/10/2015
		Lin. No.: 00022
		Soln. Date: 07/27/2015
		Soln. No.: 00179
		File Date: 07/27/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 1268

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:28D		
Control 1 EC	0.100%	08:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:30D		
Control 2 EC	0.099%	08:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:31D		
Control 3 EC	0.099%	08:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: TARR

First Name: MARYLOUISE

MI: -

Signature: \_\_\_\_\_



Badge No.: 431

Date: 07/27/2015