

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01316	Calib. Date: 03/28/2016
Certification File No.:	01317	Calib. No.: 00037
Linearity File No.:	01318	Cert. Date: 03/28/2016
Solution File No.:	01350	Cert. No.: 00024
Sequential File No.:	01350	Lin. Date: 03/28/2016
		Lin. No.: 00024
		Soln. Date: 06/23/2016
		Soln. No.: 00198
		File Date: 06/23/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	201603L	Expires: 03/22/2018
		Bottle No.: 0865

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:23D		
Control 1 EC	0.100%	13:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:24D		
Control 2 EC	0.101%	13:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:25D		
Control 3 EC	0.101%	13:26D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MOHR

First Name: DANNY

MI:

Signature: \_\_\_\_\_



Badge No.: 471

Date: 06/23/2016