

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01377	Calib. Date:	09/22/2016
Certification File No.:	01378	Cert. No.:	00038
Linearity File No.:	01379	Lin. Date:	09/22/2016
Solution File No.:	01395	Cert. No.:	00025
Sequential File No.:	01395	Lin. No.:	00025
		Soln. No.:	00206
		File Date:	12/09/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	0488

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:44S		
Control 1 EC	0.103%	08:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	08:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:45S		
Control 2 EC	0.103%	08:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	08:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46S		
Control 3 EC	0.102%	08:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	08:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MOHR

First Name: DANNY

MI:

Signature: _____



Badge No.: 471

Date: 12/09/2016

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01377	Calib. Date:	09/22/2016
Certification File No.:	01378	Cert. Date:	09/22/2016
Linearity File No.:	01379	Lin. Date:	09/22/2016
Solution File No.:	01394	Soln. Date:	12/09/2016
Sequential File No.:	01394	File Date:	12/09/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	05/17/2018
Solution Control Lot:	201605D			Bottle No.:	0484

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:26S		
Control 1 EC	-.---%	06:27S	22.9°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	06:27S	22.9°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MOHR

First Name: DANNY

MI:

Signature: _____



Badge No.: 471

Date: 12/09/2016