

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01413 Calib. Date: 02/22/2017 Calib. No.: 00040
Certification File No.: 01414 Cert. Date: 02/22/2017 Cert. No.: 00027
Linearity File No.: 01415 Lin. Date: 02/22/2017 Lin. No.: 00026
Solution File No.: 01433 Soln. Date: 05/19/2017 Soln. No.: 00212
Sequential File No.: 01433 File Date: 05/19/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 10/28/2018
Solution Control Lot: 16330 Bottle No.: 0588

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:51D		
Control 1 EC	0.099%	08:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52D		
Control 2 EC	0.099%	08:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53D		
Control 3 EC	0.099%	08:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MOHR

First Name: DANNY

MI:

Signature: _____

Badge No.: 471

Date: 05/19/2017