

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01445	Calib. Date: 07/26/2017
Certification File No.:	01446	Calib. No.: 00041
Linearity File No.:	01447	Cert. Date: 07/26/2017
Solution File No.:	01459	Cert. No.: 00028
Sequential File No.:	01459	Lin. Date: 07/26/2017
		Lin. No.: 00027
		Soln. Date: 10/14/2017
		Soln. No.: 00220
		File Date: 10/14/2017

Calibrating Unit:	4	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	17200	Expires: 06/28/2019
		Bottle No.: 0823

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:16D		
Control 1 EC	-.---%	08:16D	35.9°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	08:16D	35.9°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: _____



Badge No.: 585

Date: 10/14/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01445	Calib. Date: 07/26/2017
Certification File No.:	01446	Cert. No.: 00041
Linearity File No.:	01447	Cert. Date: 07/26/2017
Solution File No.:	01460	Lin. Date: 07/26/2017
Sequential File No.:	01460	Lin. No.: 00027
		Soln. Date: 10/14/2017
		Soln. No.: 00221
		File Date: 10/14/2017
Calibrating Unit:	4	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	17200	Expires: 06/28/2019
		Bottle No.: 0823

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:51D		
Control 1 EC	-.---%	09:51D	29.6°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	09:51D	29.6°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: 

Badge No.: 585

Date: 10/14/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01445	Calib. Date: 07/26/2017
Certification File No.:	01446	Calib. No.: 00041
Linearity File No.:	01447	Cert. Date: 07/26/2017
Solution File No.:	01461	Cert. No.: 00028
Sequential File No.:	01461	Lin. Date: 07/26/2017
		Lin. No.: 00027
		Soln. Date: 10/14/2017
		Soln. No.: 00222
		File Date: 10/14/2017

Calibrating Unit:	4	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 06/28/2019
Solution Control Lot:	17200		Bottle No.: 0823

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:57D		
Control 1 EC	-.---%	10:58D	27.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	10:58D	27.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: _____



Badge No.: 585

Date: 10/14/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01445	Calib. Date: 07/26/2017
Certification File No.:	01446	Calib. No.: 00041
Linearity File No.:	01447	Cert. Date: 07/26/2017
Solution File No.:	01462	Cert. No.: 00028
Sequential File No.:	01462	Lin. Date: 07/26/2017
		Lin. No.: 00027
		Soln. Date: 10/16/2017
		Soln. No.: 00223
		File Date: 10/16/2017
Calibrating Unit:	4	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	17200	Expires: 06/28/2019
		Bottle No.: 0823

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:34D		
Control 1 EC	0.102%	10:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:35D		
Control 2 EC	0.102%	10:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36D		
Control 3 EC	0.101%	10:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	10:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BAL

First Name: FRANK

MI: N

Signature:

Badge No.: 467

Date: 10/16/2017